



STATE OF DELAWARE  
DEPARTMENT OF CORRECTION  
245 MCKEE ROAD  
DOVER, DELAWARE 19904  
Telephone: (302) 857-5412  
Fax: (302) 622-4461

TO: ALL OFFERORS

FROM: ERIKA MARTINE-DUQUETTE  
PURCHASING SERVICES ADMINISTRATOR

SUBJECT: Addendum to RFP Contract No. DOC1213-URINALYSIS, Employee Urinalysis  
Screening Tests

**ADDENDUM #2**  
November 8, 2012

This addendum is issued to notify all interested vendors of the following change to Contract No. DOC1213-URINALYSIS, EMPLOYEE URINALYSIS SCREENING TESTS:

Pages 3 and 4 - Section II, Paragraph C, Bullet a. is revised as follows (revisions noted in red):

**C. TESTING METHODOLOGIES**

- a. Initial screening shall be done by Enzyme Multiplied Immunoassay Technique (EMIT). The immunoassays will be grouped as follows and the cutoff levels will be in accordance with Federal Health and Human Services Standards. As of **October 1, 2010**, those standards are:

| <u>Drug</u>           | <u>ng/ml</u> |
|-----------------------|--------------|
| Marijuana metabolites | 50           |
| Cocaine metabolities  | <b>150</b>   |
| Phencyclidine (PCP)   | 25           |
| Opiate metabolites    | 2000         |
| Amphetamines          | <b>500</b>   |

All other terms and conditions remain the same.

## Analytes and Their Cutoffs

Effective Date: October 1, 2010

Reference: Federal Register, November 25, 2008 (73 FR 71858), Section 3.4

| <b>Initial test analyte</b>                         | <b>Initial test cutoff concentration</b> | <b>Confirmatory test analyte</b>              | <b>Confirmatory test cutoff concentration</b> |
|---|--|---|---|
| Marijuana metabolites                               | 50 ng/mL                                 | THCA <sup>1</sup>                             | 15 ng/mL                                      |
| Cocaine metabolites                                 | 150 ng/mL                                | Benzoyllecgonine                              | 100 ng/mL                                     |
| Opiate metabolites<br>Codeine/Morphine <sup>2</sup> | 2000 ng/mL                               | Codeine<br>Morphine                           | 2000 ng/mL<br>2000ng/mL                       |
| 6-Acetylmorphine                                    | 10 ng/mL                                 | 6-Acetylmorphine                              | 10 ng/mL                                      |
| Phencyclidine                                       | 25 ng/mL                                 | Phencyclidine                                 | 25 ng/mL                                      |
| Amphetamines <sup>3</sup><br>AMP/MAMP <sup>4</sup>  | 500 ng/mL                                | Amphetamine<br>Methamphetamine <sup>5</sup>   | 250 ng/mL<br>250 ng/mL                        |
| MDMA <sup>6</sup>                                   | 500 ng/mL                                | MDMA<br>MDA <sup>7</sup><br>MDEA <sup>8</sup> | 250 ng/mL<br>250 ng/mL<br>250 ng/mL           |

<sup>1</sup> Delta-9-tetrahydrocannabinol-9-carboxylic acid (THCA).

<sup>2</sup> Morphine is the target analyte for codeine/morphine testing.

<sup>3</sup> Either a single initial test kit or multiple initial test kits may be used provided the single test kit detects each target analyte independently at the specified cutoff.

<sup>4</sup> Methamphetamine is the target analyte for amphetamine/methamphetamine testing.

<sup>5</sup> To be reported as positive for methamphetamine, a specimen must also contain amphetamine at a concentration equal to or greater than 100 ng/mL.

<sup>6</sup> Methylenedioxyamphetamine (MDMA).

<sup>7</sup> Methylenedioxyamphetamine (MDA).

<sup>8</sup> Methylenedioxyethylamphetamine (MDEA).